

Troubleshooting Manual

Model 3370 Air Cooled Recirculator

Table of contents

Section 1. Service and Technical Support

Section 2. Replacement Parts

Section 1. Service and Technical Support

If your recirculator fails to operate properly, contact the location where the unit was purchased. Have the following information available for customer service:

- Model and Serial Number
- Voltage (from back panel label)
- Date of purchase and your purchase order number
- Suppliers' order number or invoice number
- A summary of the problem

Section 2. Replacement Parts

Pump Motor (120V, 60Hz and 240V, 50Hz)	215-217
Pump Head (2.3 gpm)	215-218
Air Filter, Reusable	400-369
Water Filter Assembly, Nylon, with Screen	300-098
Fan Motor, 120V, 60Hz	215-194
Fan Motor, 240V, 50Hz	215-188
Fan Blade	215-076
PolyScience Lab Algicide	4-3000340