

# Certificate



Certificate no.

CU 72150294 01

**License Holder:**  
PolyScience  
6600 West Touhy Avenue  
Niles IL 60714-4516  
USA

**Manufacturing Plant:**  
PolyScience  
6600 West Touhy Avenue  
Niles IL 60714-4516  
USA

**Test report no.:** USA-LG 31570323 002      **Client Reference:** Melanie Nagy  
**Tested to:** UL 61010-1:2012  
UL 61010-2-010:2015  
CAN/CSA-C22.2 NO. 61010-1-12  
CAN/CSA-C22.2 NO. 61010-2-010-04 (R2014)

**Certified Product:** Liquid to Air Recirculator

**License Fee - Units**

**Model Designation:**

- 1) 1/4 to 1HP, 3 Series: 3XXXXXXXXXX;
  - 2) 1/4 to 1HP, 5 Series: 5XXXXXXXXXX;
  - 3) 1/4 to 1HP, 6 Series: 6XXXXXXXXXX;
  - 4) 1.5HP, 5 Series: 5XXXXXXXXXX;
  - 5) 1.5HP, 6 Series: 6XXXXXXXXXX;
  - 6) 3370 Series: 337XXXXXXXX;
  - 7) 4000 Series: 4XXXXXXXXXX;
  - 8) L, M, H Series: LXXXXXXXXXX, MXXXXXXXXXX,  
HXXXXXXXXXX
- (X = 0-9, A-Z; see Appendix (Constr. Data Form))

7

contd.

7

Appendix: 1, 1-25

**Licensed Test mark:**



**Date of Issue**  
(day/mo/yr)  
05/08/2015

# Certificate



Certificate no.

CU 72150294 02

**License Holder:**  
PolyScience  
6600 West Touhy Avenue  
Niles IL 60714-4516  
USA

**Manufacturing Plant:**  
PolyScience  
6600 West Touhy Avenue  
Niles IL 60714-4516  
USA

**Test report no.:** USA-LG 31570323 002

**Client Reference:** Melanie Nagy

**Tested to:** UL 61010-1:2012  
UL 61010-2-010:2015  
CAN/CSA-C22.2 NO. 61010-1-12  
CAN/CSA-C22.2 NO. 61010-2-010-04 (R2014)

**Certified Product:** Liquid to Air Recirculator

**License Fee - Units**

contd.

**Rated Voltage:**

- 1) AC 120V/230V/240V, 60/60/50Hz
- 2) AC 120V/230V/240V, 60/(50/60)/60/50Hz
- 3) AC 120V/230V/240V/460V, 60/60/50/50Hz
- 4) 3 AC 208-230V/230V/240V/380V, 50/60Hz
- 5) 3 AC 230V/240V/460V, 50/60Hz
- 6) AC 120V, 50/60Hz; AC 230V/240V, 50/60Hz
- 7) AC 120V/230V, 50/60Hz
- 8) AC 120V/230V/240V, 60/60/50Hz

contd.

**Licensed Test mark:**



**Date of Issue**  
(day/mo/yr)  
05/08/2015

# Certificate



Certificate no.

CU 72150294 03

**License Holder:**  
PolyScience  
6600 West Touhy Avenue  
Niles IL 60714-4516  
USA

**Manufacturing Plant:**  
PolyScience  
6600 West Touhy Avenue  
Niles IL 60714-4516  
USA

**Test report no.:** USA-LG 31570323 002

**Client Reference:** Melanie Nagy

**Tested to:** UL 61010-1:2012  
UL 61010-2-010:2015  
CAN/CSA-C22.2 NO. 61010-1-12  
CAN/CSA-C22.2 NO. 61010-2-010-04 (R2014)

**Certified Product:** Liquid to Air Recirculator

**License Fee - Units**

contd.

**Rated Current:**

- 1) 16.0A/12.2A/12.5A max.
- 2) 13.5A/12.2A/12.5A max.
- 3) 16.0A/12.2A/12.5A/7.8A max.
- 4) 35A max.
- 5) 29A max.
- 6) 8A; 4A
- 7) 7.0A max.
- 8) 12A/6A/6A max.

**Protection Class:** I

**Licensed Test mark:**



**Date of Issue**  
**(day/mo/yr)**  
05/08/2015